




Supplementary Table S1. Comparison between MitraClip and PASCAL.

	MitraClip	PASCAL
Device	 NT NTW XT XTW	 ACE P10
Width	NT and XT – 4 mm NTW and XTW – 6 mm	P10 – 10 mm Ace – 6 mm
Retention Elements	NT and NTW – 4 rows XT and XTW – 6 rows	Single row
Spacer	No	Yes
Independent Grasping	Yes	Yes
Material	Cobalt Chromium	Nitinol
Locking Mechanism	Active	Passive
Delivery System	 2 Catheters Steerable Guide Catheter and Clip Delivery System	 3 Catheters Guide Sheath, Steerable Catheter and Implant Catheter